

CERTIFICATE OF CALIBRATION

Certificate #: 90052

Keep for your records.

<u>Customer:</u>	Sample Customer 123 Any Street Any Town, CA 12345 United States	Laboratory Location:	Max Machinery, Inc 33A Healdsburg Ave Healdsburg, CA 95448 707-433-2662				
<u>Type of Device:</u>	Flow Meter	<u>Calibration Fluid:</u>	Hydra Oil				
<u>Manufacturer:</u>	Max Machinery, Inc.	<u>Fluid Viscosity:</u>	30 cps				
<u>Model Number:</u>	G004HS13NA/A11A/1	<u>Fluid Specific Gravity:</u>	0.85 g/mL				
<u>Serial Number:</u>	D12345	<u>Fluid Temp:</u>	21°C +/- 1°C				
Date of Calibration:	7/1/2015	<u>Output Units:</u>	mA				
Sales Order:	SAMPLE	<u>Flow Units:</u>	mL/min				
Procedure Used:	LA-P-107	<u>Max Output:</u>	20 mA @ 200 mL/min				
Performed By:	JDO	<u>Min Output:</u>	4 mA @ 0 mL/min				

Calibration Notes:

This document reflects the new linear calibration. The new condition was found to be in tolerance.

Calibration Data

Flow Rate	Output	Error		Flow Rate	Output	Error -	
mL/min	Milliamps	% reading	mA Offset	mL/min	Milliamps	% reading	mA Offset
191.70	19.341	0.03%			-		
150.83	16.069	0.02%			TEL		
99.13	11.931	0.01%			01		
73.66	9.893	0.00%	LADDD	ATOPY	L E		
38.83	7.105	-0.05%	LADON	aloni	10		
18.01	5.439		-0.002		100		
		100	1		11		
		- A	0		14		
		1	Contraction of the second				
			200				

Errors for flow rates below 10% of the full scale output are represented as a milliamp offset, rather than a percentage.

Equipment Used in the Calibration:

Calibration ID:	Description:	Serial Number:	Cal Due Date:	Certificate Number:
41901	Built in reference meter	C740722	3/31/2016	41901033115
41902	Built in reference meter	C740745	3/31/2016	41902032615
41903	Built in reference meter	C740752	3/31/2016	41903041315
41904	Temperature Controller	28133	11/11/2015	41904111214
41906	Multifunction DAQ	16769D4	11/26/2015	41906112614
41907	Counter/timer	1677AE2	11/26/2015	41907112614

QC Approval:

ne Doc 7/1/2015

Calibration Technician

Lab Technician

7/1/2015 John Doe

Jane Doe

Quality Manager

This calibration was conducted using standards traceable to NIST.

Measurement uncertainty of the #419 test stand is +/- 0.167% of reading with a 95% confidence (k=2 coverage factor).

Calculations are available upon request.

This Certificate shall not be reproduced, except in full, without written approval by Max Machinery, Inc.

33A Healdsburg AvenueTHealdsburg, CA 95448F

T +1 707.433.2662 F +1 707.433.1818 Print Date: 23 Nov 2016 1:12:44 PM Page 1 of 1