



Max Machinery, Inc.
an ISO 9001:2008 certified company

CERTIFICATE OF CALIBRATION

Certificate #: 90055

Keep for your records.

Customer: Sample Customer
123 Any Street
Any Town, CA 12345
United States

Laboratory Location: Max Machinery, Inc
33A Healdsburg Ave
Healdsburg, CA 95448
707-433-2662

Type of Device: Flow Meter
Manufacturer: Max Machinery, Inc.
Model Number: G045HS11NA/P11N/1
Serial Number: D12345

Calibration Fluid: Hydra Oil
Fluid Viscosity: 30 cps
Fluid Specific Gravity: 0.85 g/mL
Fluid Temp: 21°C +/- 1°C

Date of Calibration: 7/1/2015
Sales Order: SAMPLE
Procedure Used: LA-P-107
Performed By: JDO

Output Units: Pulses/L
Flow Units: L/min
K-Factor: 70000

Calibration Notes: This document reflects the new linear calibration.
The new condition was found to be in tolerance.

Calibration Data

Flow Rate	----- Output -----		Error	Flow Rate	----- Output -----		Error
L/min	Pulses/L	Frequency Hz	% reading	L/min	Pulses/L	Frequency Hz	% reading
45.00	70011	52508.250	0.02%				
35.00	70019	40844.420	0.03%				
15.00	69997	17499.250	0.00%				
5.00	70047	5837.250	0.07%				
1.05	70020	1225.350	0.03%				
0.45	70019	525.143	0.03%				
0.20	70030	233.433	0.04%				

Equipment Used in the Calibration:

Calibration ID:	Description:	Serial Number:	Cal Due Date:	Certificate Number:
41901	Built in reference meter	C740722	3/31/2016	41901033115
41902	Built in reference meter	C740745	3/31/2016	41902032615
41903	Built in reference meter	C740752	3/31/2016	41903041315
41904	Temperature Controller	28133	11/11/2015	41904111214
41906	Multifunction DAQ	16769D4	11/26/2015	41906112614
41907	Counter/timer	1677AE2	11/26/2015	41907112614

QC Approval:

Jane Doe

Jane Doe
Quality Manager
7/1/2015

Calibration Technician

John Doe

John Doe
Lab Technician
7/1/2015

This calibration was conducted using standards traceable to NIST.
Measurement uncertainty of the #419 test stand is +/- 0.167% of reading with a 95% confidence (k=2 coverage factor).
Calculations are available upon request.

This Certificate shall not be reproduced, except in full, without written approval by Max Machinery, Inc.