

## **CERTIFICATE OF CALIBRATION**

Certificate #: 90049

Keep for your records.

<u>Customer:</u>	Sample Customer 123 Any Street Any Town, CA 12345 United States	Laboratory Location:	Max Machinery, Inc 33A Healdsburg Ave Healdsburg, CA 95448 707-433-2662
Type of Device: Manufacturer: Model Number: Serial Number:	Flow Meter Max Machinery, Inc. H241LS11NA/P11N/0 D12345	<u>Calibration Fluid:</u> <u>Fluid Viscosity:</u> <u>Fluid Specific Gravity:</u> Fluid Temp:	Hydra Oil 30 cps 0.85 g/mL 21°C +/- 1°C
Date of Calibration: Sales Order: Procedure Used: Performed By:	7/1/2015 SAMPLE LA-P-412 JDO	<u>Output Units:</u> Flow Units: <u>K-Factor:</u>	Pulses/L L/min 15000

<u>Calibration Notes:</u> This document reflects the new linear calibration. The new condition was found to be in tolerance.

## **Calibration Data**

Out	put	Error	Flow Rate	Outp	out	Error
Pulses/L	Frequency Hz	% reading	L/min	Pulses/L	Frequency Hz	% reading
15002	45006.000	0.01%		-		
15002	25003.330	0.01%		TEL		
15001	15001.000	0.01%		0		
15002	5000.667	0.01%	ATOPY			
15000	2500.000	0.00%	anoni j			
14994	1249.500	-0.04%	10	100		
14999	499.967	-0.01%		TIN .		
		200	_			
		0.01				
	Pulses/L 15002 15002 15001 15002 15000 14994	15002  45006.000    15002  25003.330    15001  15001.000    15002  5000.667    15000  2500.000    14994  1249.500	Pulses/L  Frequency Hz  % reading    15002  45006.000  0.01%    15002  25003.330  0.01%    15001  15001.000  0.01%    15002  5000.667  0.01%    15000  2500.000  0.00%    14994  1249.500  -0.04%	Pulses/L  Frequency Hz  % reading  L/min    15002  45006.000  0.01%	Pulses/L  Frequency Hz  % reading  L/min  Pulses/L    15002  45006.000  0.01% <td< td=""><td>Pulses/L  Frequency Hz  % reading  L/min  Pulses/L  Frequency Hz    15002  45006.000  0.01%</td></td<>	Pulses/L  Frequency Hz  % reading  L/min  Pulses/L  Frequency Hz    15002  45006.000  0.01%

## **Equipment Used in the Calibration:**

Calibration ID:	Description:
41201	Built in reference meter
41202	Built in reference meter
41203	Temperature controller
41206	Multifunction DAQ
41207	Counter/timer

Serial Number: B407001 B407007 7084 15CD165 165894

## Cal Due Date: 8/31/2015 8/31/2015 11/12/2015 11/26/2015 11/26/2015

Certificate Number: 41200083112 41201083112 41203111214 41206112614 41207112614

QC Approval:

Jane Doe

7/1/2015

**Calibration Technician** 

7/1/2015

John Doe Lab Technician

Quality Manager

This calibration was conducted using standards traceable to NIST.

Measurement uncertainty of the #412 test stand is +/- 0.113% of reading with a 95% confidence (k=2 coverage factor). Calculations are available upon request.

This Certificate shall not be reproduced, except in full, without written approval by Max Machinery, Inc.

Max Machinery, Inc. maxmachinery.com 33A Healdsburg AvenueTHealdsburg, CA 95448F

T +1 707.433.2662 F +1 707.433.1818 Print Date: 23 Nov 2016 1:11:24 PM Page 1 of 1