

## **CERTIFICATE OF CALIBRATION**

Certificate #: 90059

Keep for your records.

Sample Customer **Customer:** 

> 123 Any Street Any Town, CA 12345

**United States** 

**Laboratory Location:** Max Machinery, Inc

**Calibration Fluid:** 

**Fluid Specific Gravity:** 

Fluid Viscosity:

Fluid Temp:

**Output Units:** 

**Flow Units:** 

K-Factor:

33A Healdsburg Ave Healdsburg, CA 95448

707-433-2662

Kerosene

0.82 g/mL

Pulses/mL

mL/min

1000.0

21°C +/- 1°C

3 cps

Type of Device: Flow Meter

Manufacturer: Max Machinery, Inc. P002HS25NA/P11N/1 **Model Number:** 

**Serial Number:** D12345

7/1/2015 **Date of Calibration:** Sales Order: **SAMPLE** LA-P-110 **Procedure Used:** JDO Performed By:

**Calibration Notes:** 

This document reflects the new linear calibration.

The new condition was found to be in tolerance.

**Calibration Data** 

Flow Rate	Ou	Error	
mL/min	Pulses/mL	Frequency Hz	% reading
2000.00	1000.6	33353.330	0.06%
1800.00	1000.6	30018.000	0.06%
1500.00	1000.6	25015.000	0.06%
1000.00	1000.4	16673.330	0.04%
300.00	999.9	4999.500	-0.01%
100.00	999.9	1666.500	-0.01%
30.00	999.9	499.950	-0.01%
10.00	999.7	166.617	-0.03%
3.00	999.7	49.985	-0.03%
1.00	1001.7	16.695	0.17%

Flow Rate	Outp	out	Error	
mL/min	Pulses/mL	Frequency Hz	% reading	
N W 10				
	The state of the s			
20				
TODY	MANUAL PROPERTY.			
HOKY	_6_			
2/0	7/2			
	500			
	A			
1000				

## **Equipment Used in the Calibration:**

<b>Calibration ID:</b>	Description:	Serial Number:	Cal Due Date:	Certificate Number:
41701	Built in reference meter	D22136	3/6/2016	41701030615
41702	Built in reference meter	C740723	3/6/2016	41702022715
41703	Temperature controller	2430MR	11/12/2015	41703111214
41704	Counter/timer	166FC42	11/26/2015	41704112614
41706	Multifunction DAQ	16769FA	11/16/2015	41706112614

QC Approval:

**Calibration Technician** 

Jane Doe

7/1/2015

7/1/2015

**Quality Manager** 

John Doe

Lab Technician

This calibration was conducted using standards traceable to NIST. Measurement uncertainty of the #417 test stand is +/- 0.176% of reading with a 95% confidence (k=2 coverage factor). Calculations are available upon request.

This Certificate shall not be reproduced, except in full, without written approval by Max Machinery, Inc.

Print Date: 23 Nov 2016 1:14:24 PM

Page 1 of 1